




Spectral emissivity measurement



Laser heated sample





Spectral range: 200nm – 20 μm
Temperature: -80 - 1600°C




Small IR sources

Dynamic sample heating




Actions: IR thermometer replaced by custom made IR sensor (TP with KRS-5 window), sample holder, "wall" for rotated sample


New EMIRIM DTU SETUP



Time slope - ΔT





DTU temperature ramp method




Results for 3 samples. Emissivity of Tesa 50525 is found to 0.0178, Foil (no identification, Mylar type coated film) $\epsilon=0.2866$, Rockw 11.4g $\epsilon=0.024$. Uncertainty of measurements can be improved by replacing thermocouple with a calibrated pt100 temperature sensor and better temperature control of the instrumentation and room. Results for 8-14 μm band.

Enclosure & IR-sensor



- 345x345x345 mm³ cube
- 1.8° step angle sample
- Peltier cooling/heating of sample...
- 1 mK Peltier stabilized IR-sensor KRS-window
- Low cost using commercial parts



DTU setup

